

Preparation of antisera for serological detection of streptococcal pneumoniae antigens.*

Kamjorn Tatiyakavee**

Mai Ratanavararuk** Pongpun Nanthapisud**

Reutai Sakulramrung** Nipa Rujithamkul**

Tatiyakavee K, Ratanavararuk M, Nanthapisud P, Sakulramrung R, Rujithamkul N. Preparation of antisera for serological detection of streptococcal pneumoniae antigens. Chula Med J 1990 July; 33 (7): 501-508

Streptococcus pneumoniae is an important etiologic agent of lower respiratory tract infection. Serological detection of Streptococcus pneumoniae antigens from clinical specimens is a useful diagnostic method. In this study, we have prepared antisera against the antigens of Streptococcus pneumoniae type 1, 19, 5, 6, and 23, which are the common serotypes found in Thailand. The reference strains of Streptococcus pneumoniae were kindly provided from Streptococcal Department, Statens Seruminstitut, Denmark. The antisera were prepared by using 5 rabbits for each strain with at least 5 weeks period of immunization.

The geometric mean titers of antisera against Streptococcus pneumoniae type 1, 19, 5, 6, and 23 were 194, 2435.5, 1024, 55, and 194 respectively. There were cross-reactions between type 1 antiserum and type 6 antigen, type 6 antiserum and type 1 antigen, type 6 antiserum and Hemophilus influenzae antigen, antisera of type 1, 6, 23 and Escherichia coli. Respective immune immunoglobulins were purified by affinity chromatography on a Protein A-sepharose CL-4B column and coupled to latex particles for testing against various heterologous bacteria from clinical specimen. The preliminary testing of this latex agglutination kit showed good result without cross-reaction.

Reprint request : Tatiyakavee K, Department of Microbiology, Faculty of Medicine, Chulalongkorn University, Bangkok 10330, Thailand.

Received for publication. May 2, 1990.

* This study received support from the Rachadapisaksompoj-China Medical Board Research Funds.

** Department of Microbiology, Faculty of Medicine, Chulalongkorn University.

ก่าจร คติยกวี, โหม รัตนวราธิ์, ผ่องพรรณ นันทาภิสูทธิ, ฤทธิ สกูลแรมรุ่ง, นิภา ฤทธิธรรมกุล. การเตรียมแอนติเซรา ต่อแอนติเจนของ สเตร็ปโตคอคคาล นิโวเนีย. จุฬาลงกรณ์เวชสาร 2533 กรกฎาคม ; 33 (7): 501-508

Streptococcus pneumoniae เป็นจุลชีพสำคัญที่ทำให้เกิดโรคติดเชื้อของระบบหายใจส่วนล่าง การตรวจหาแอนติเจนของ *Streptococcus pneumoniae* จากสิ่งส่งตรวจต่าง ๆ โดยวิธีทาง serology เป็นอีกวิธีหนึ่งซึ่งช่วยการวินิจฉัยโรคติดเชื้อที่เกิดจากจุลชีพดังกล่าว และเป็นวิธีที่มีประโยชน์อย่างยิ่งในกรณีที่มีการเพาะเชื้อไม่ได้ผล งานวิจัยนี้เป็นการเตรียมแอนติเซรัมเพื่อให้ในการตรวจหาแอนติเจนของ *Streptococcus pneumoniae* โดยเลือกสายพันธุ์ที่ทำให้เกิดโรคติดเชื้อ ได้บ่อยจากรายงานในประเทศไทยเพียง 5 สายพันธุ์คือ 1, 19, 5, 6 และ 23 ทำการเตรียมวัคซีนสำหรับฉีดกระต่ายด้วยเชื้อมาตรฐานที่ใช้อย่างซึ่งได้รับมาจาก *Streptococcal Department, Statens Seruminstitut, Denmark* ฉีดวัคซีนโดยใช้กระต่าย 5 ตัว สำหรับการเตรียมแอนติเซรัมแต่ละสายพันธุ์ การฉีดวัคซีนให้กระต่ายแต่ละชุดใช้เวลาประมาณ 5 สัปดาห์จึงจะเอาเลือดแยกแอนติเซรัม

ผลการศึกษาพบว่า การตอบสนองสร้างแอนติบอดีของกระต่ายต่อเชื้อแต่ละสายพันธุ์มีขัณมิเรขาคณิตของ Titer ของแอนติบอดีต่อสายพันธุ์ 1, 19, 5, 6 และ 23 เป็น 194, 2435.5, 1024, 55 และ 194 ตามลำดับ เมื่อนำแอนติเซรัมที่เตรียมได้ไปทดสอบหาปฏิกิริยาข้ามกลุ่มระหว่างสายพันธุ์ พบปฏิกิริยาข้ามกลุ่มกันชัดเจน ระหว่างแอนติเซรัมต่อสายพันธุ์ 1 และ 6 การทดสอบปฏิกิริยาข้ามกลุ่มกับเชื้ออื่น ๆ พบว่าแอนติเซรัมต่อสายพันธุ์ 6 มีปฏิกิริยาข้ามกลุ่มต่อ *Hemophilus influenzae* และ แอนติเซรัมต่อสายพันธุ์ 1,6,23 มีปฏิกิริยาข้ามกลุ่มต่อ *Escherichia coli* นำแอนติเซรัมไปแยกสกัดเฉพาะ *Immunoglobulins* โดยวิธี *affinity chromatography* ด้วย *Protein A sepharose CL-4B column* นำไปประยุกต์ประคิษฐ์ชุดทดสอบ *Latex agglutination* ซึ่งจากการทดสอบเบื้องต้นพบว่าชุด *Latex agglutination* ที่เตรียมขึ้นใช้เองนี้มีประสิทธิภาพดี

Communicable diseases remain the main health problem in Thailand. *Streptococcus pneumoniae* is one of the most common causes of lower respiratory tract infection especially in children under 2 years^(1,2). The potentially serious nature of the disease requires the early diagnosis of infection to ensure the institution of appropriate antimicrobial treatment. Different kinds of soluble pneumococcal capsular polysaccharide are used to produce more than 80 type-specific commercial diagnostic antisera⁽³⁾. However, some epidemiological studies in Thailand indicated that only a few types of *Streptococcus pneumoniae*, including type 1, 6, 19, 5 and 23, frequently caused infections in hospitalized patients⁽⁴⁾.

At present Counter-Immunoelectrophoresis (CIE)⁽⁵⁻⁷⁾ and latex agglutination (LA) are two of the most acceptable serological methods for detection of *Streptococcus pneumoniae* from clinical specimens.⁽⁸⁻¹⁰⁾ However, the performing of CIE requires electrophoretic equipments and uses more standard antisera than the LA method. In addition, the LA method is easy to interpret and suitable for bed-side diagnosis^(11,12).

In this paper, we describe a simple method in the preparation of specific antisera against *Streptococcus pneumoniae* serotypes 1, 6, 19, 5 and 23. The aim of the present investigation is to evaluate the prepared antisera compared with the commercial antisera. Furthermore it is to use the prepared antisera with the appropriate serological method, latex agglutination, for *Streptococcus pneumoniae* screening.

Materials and Methods

1. Lyophilized standard strains of *Streptococcus pneumoniae* serotypes 1, 5, 6, 19 and 23 (kindly provided from Streptococcal Department, Staten Serum Institut, Copenhagen, Denmark) were used for vaccine preparation.

2. Standard commercial antisera for serotypes 1, 5, 6, 19 and 23 were obtained from Staten Serum Institut, Copenhagen, Denmark.

3. White New Zealand rabbits, weighing 2-2.5 kilogram, were used for immunizations.

Vaccine preparation

Vaccine preparations followed Lund's method⁽¹³⁾. Each selected strain of lyophilized *Streptococcus pneumoniae* was dissolved in brain heart infusion broth and then added to sheep blood agar. The culture was incubated at 37°C in 5% CO₂ for 24 hours; this was a control of their sterility. Then, a single isolated colony from the pure culture was inoculated into 10 ml of 5% horse serum broth and incubated at 37°C for 18 hours. In the morning the serum-broth culture was expanded into

a flask of 1500 ml of 5% horse serum broth. The flask was shaken and incubated again at 37°C, and the growth was sampled by culture in sheep blood agar every hour.

Gram stain was performed to exclude impurity. The density of the growth in each sample was estimated, and the capsule formation is tested by Quellung's reaction,⁽¹⁴⁾ by its specific antiserum. The optimal density was reached after about 6 hours; then the growth was stopped by the addition of 2% formalin. The flask remained in the incubator at 37°C until the following morning. The culture was then centrifuged 3000 g for 30 min. The sediment was suspended in 0.5% formalin in Sorensen buffer solution (20 ml of M/15 KH₂PO₄ + 80 ml of M/15 Na₂HPO₄ · 2H₂O + 300 ml of 0.9% NaCl solution). The density of the finished vaccine was estimated by means of a standard containing 1.2 × 10⁹ organisms per ml. The vaccine was divided into several small tubes and stored at -30°C.

Immunizations

Each strain of *Streptococcus pneumoniae* needed 5 white rabbits for immunization. At first, blood samples of all rabbits were taken, and then their sera were poured and frozen. The vaccine was warmed up to 37°C prior to the injections, which was given slowly and intravenously into their ears. The rabbits were injected with increasing dosages of vaccine every other day (three times a week) during a period of 5 weeks. The schedule of immunizations was arranged as in the following table (Table 1).

Production of antisera

After 5 weeks of immunization, blood samples were taken by venepuncture to estimate the potency of the antisera. The various rabbits did not produce the same amounts of antibody. Further immunization was given by repeating the 5th week's dosage scheme of the schedule (Table 1) in the rabbits that produced low titre of antibody. When the sera were found to possess sufficiently high titre, the rabbits were bled by venepuncture or heart puncture to take at least 30 ml of blood from each rabbit.

The blood samples were left standing in cold-room for 24 hours. After centrifugation, 1% thimerosal was added to the obtained sera and stored at -30°C.

All sera were examined for agglutination titre with microagglutination method, and tested for their specificities with Quellung's reaction.

Microagglutination method⁽¹⁵⁾

The prepared antisera were serially two fold diluted with 0.85% saline in wells of microtitre plates. The type-specific *Streptococcus pneumoniae* vaccines used

as antigens were added to the correspondent wells (Table 2). One to two drops of 10% methylene blue were added

to enhance interpretations. The positive antigen-antibody reactions showed agglutinations at the bottom of the wells.

Table 1. The schedule of immunization within 5 weeks.

	day 1	day 3	day 5
week 1	0.25 ml	0.5 ml	0.5 ml
week 2	1.0 ml	1.0 ml	1.0 ml
week 3	1.5 ml	1.5 ml	1.5 ml
week 4	2.0 ml	2.0 ml	2.0 ml
week 5	2.0 ml	2.0 ml	2.0 ml

Table 2. Titration by microagglutination procedure in microtiter plate.

	well 1	well 2	well 3	well 4	well 5	well 6	well 7	well 8	well 9
0.85% saline antibody	<-----50 μ l----->								50 μ l
Ab* titer	1	2	4	8	16	32	64	128	-ve
antigen	<-----50 μ l----->								50 μ l

Ab* : antibody

Immunoglobulin G extraction for latex agglutination

Immunoglobulin G was extracted from the prepared antisera by affinity chromatography method. Protein A coupling to sepharose CL-4B was used to bind IgG molecules from the applied sera⁽¹⁶⁾. Protein A is a protein isolated from the cell wall of some strains of staphylococcus aureus that specifically reacts with IgG molecules of subclass IgG1, IgG2 and IgG4. The column of Protein-A sepharose CL-4B saturated in Tris-NaCl at pH 8.6, adsorbed IgG from the sera. Then the unbound substances was washed out of the column; and the purified IgG was recovered by changing the pH by acetate-NaCl pH 4.3 buffer. The concentrations of the eluted protein

were determined by optical density at 280 nm (OD₂₈₀). The purity of IgG was tested with anti-rabbit IgG and anti-rabbit serum by immunodiffusion technique.

The protein in extracted solutions were concentrated by Amicon Ultrafiltration of membrane size YM 100. The final volumes were the same as the beginning volumes of antisera before extractions. The concentrations of IgG were determined by Lowry method⁽¹⁷⁾.

Results

1. Titres of the prepared antisera before extractions of IgG were determined by microagglutination method. Geometric mean of each antisera was calculated with the following formula :

Geometric mean = n^{th} root of (titer₁ × titer₂ × titer₃ ... × titer_n) (titer_{1,2,3...n} is the agglutination titers of a type-specific antisera from the

first, second, third immunized rabbits and so on.) The geometric means of type 1, 5, 6, 19, and 23 were 194, 2435.5, 55, 1024, and 194 respectively (Table 3).

Table 3. Titer of the prepared Streptococcus pneumoniae antisera.

Antisera	range	Geometric mean
Pn* 1	64-512	194
Pn 5	512-8192	2435.5
Pn 6	8-512	55
Pn 19	256-16384	1024
Pn 23	128-256	194

Pn* = Streptococcus pneumoniae

2. Cross-reactions to the different serotypes and some of the different organisms, particularly Hemophilus influenzae type b, Escherichia coli, and Klebsiella species, were done by microagglutination method. The most cross-reactions were caused by the antiserum of type 6. Other cross-reactions are showed in table 4.

3. The same type antisera were pooled together. The pooled antisera were found to react specifically to their original vaccines by Quellung's reaction.

4. IgG extractions by affinity chromatography were done after absorption of the antisera with the cross-reacted E.coli. The titers of the antisera type 1, 5, 6, 19, and 23 were decreased to 64, 128, 256, 512 and 64 respectively (Table 5).

Table 4. Cross-reactions against other Streptococcus pneumoniae and other organisms.

Antigen	Antisera				
	Pn1	Pn5	Pn6	Pn19	Pn23
Other Pn*					
Pn 1	+	-	+	-	-
Pn 5	-	+	-	-	-
Pn 6	+	-	+	+	-
Pn 19	-	-	+	+	-
Pn 23	-	-	-	±@	+
Other bacteria					
H. influenzae b	-	-	+	-	-
E.coli	+	±@	+	-	+
Klebsiellae sp.	-	-	-	-	-

Pn* = Streptococcus pneumoniae

±@ = 2 from 5 antisera caused cross-reactions

Table 5. Comparison of antisera titer before and after IgG extractions.

Antisera	before extraction	after extraction
Pn 1	256	64
Pn 5	1024	128
Pn 6	512	256
Pn 19	2048	512
Pn 23	256	64

Pn* = *Streptococcus pneumoniae*

5. The extracted antisera were filtrated by Amicon Ultrafiltration. The final volume of each antisera was equal to its original volume before extraction, and its protein concentration was determined by Lowry method. The most concentrated antiserum was antiserum of type 19, 5.80 mg/ml, while the least was type 23, 0.68 mg/ml (Table 6).

6. Latex agglutination kit was prepared by these IgGs and tested for cross-reactions. The antisera type 5 and 19 cross-reacted to some different types of *Streptococcus pneumoniae*, but did not cross-reacted to *Hemophilus influenzae* (Table 7).

Table 6. Protein determinations for IgG by Lowry's method.

IgG of	concentration (mg/ml)
Pn* 1	1.88
Pn 5	1.26
Pn 6	3.20
Pn 19	5.80
Pn 23	0.68

Pn* = *Streptococcus pneumoniae***Table 7.** Tests for cross-reactions of Latex agglutination kit.

Antigen	Latex agglutination against				
	Pn1	Pn5	Pn6	Pn19	Pn23
Pn* 1	+	+	-	-	-
Pn 5	-	+	-	+	-
Pn 6	-	-	+	+	-
Pn 19	-	+	-	+	-
Pn 23	-	+	-	-	+
Hib #	-	-	-	-	-

Pn* = *Streptococcus pneumoniae*Hib # = *Hemophilus influenzae*

The Latex agglutination kit was tested with various clinical specimens. The preliminary result was comparable to standard culture procedure (data not shown).

Summary and discussion

Streptococcus pneumoniae antisera specific to serotype 1, 5, 6, 19, and 23 were prepared. White rabbits were immunized with vaccines prepared from the reference strains provided by Statens Serum Institut, Denmark. The antibody productions from the immunized rabbits were varied. The rabbits immunized with *Streptococcus pneumoniae* type 5 vaccine showed the highest antibody response, while immunization with *Streptococcus pneumoniae* type 6 vaccine showed poor response. However, the overall antibody responses were satisfactory. The titers' geometric means of type 1, 5, 6, 19, and 23 were 194, 2435.5, 55, 1024, and 194 respectively.

Antiserum of *Streptococcus pneumoniae* type 6 was found to have cross-reaction with *Hemophilus influenzae* type b. In addition, the antiserum also cross-reacted with *E.coli*. This is believed to be due to their antigens contain the same sugar, ribitol, in their structures (18,19) (fig. 1).

<p>H.influenzae type b ; -->3)-B-d-Rib f-(1-->1)- D-ribitol-5 phosphate (1,4) S.pneumoniae type 6 ; -->2)-L-D-Gal p-(1-->3)-L-Rhap-(1-->3)- D-ribitol-5 phosphate E.coli K100 ; Ribosyl (1-2) ribitol linkage</p>
--

Figure 1. Antigenic structures of *Streptococcus pneumoniae* type 6, *Hemophilus influenzae* type b and *Escherichia coli* K 100.

Affinity chromatography, using Protein-A sepharose CL-4B, was performed to extract IgG from the prepared antisera. The antibody titers were decreased during the extraction processes. However, the extracted IgGs were pure and suitable for coating with latex particles. Our prepared Latex agglutination kit was effective in the detection of the specific types of *Streptococcus pneumoniae* from the cultures. The preliminary testing of the kit with clinical specimens showed good result. This study is beneficial for preparations of *Streptococcus pneumoniae* antisera. Development of home made serological diagnostic kits will decrease the expenditure of using commercial antisera or kits, as

well as valuable for bed side diagnosis of certain infectious diseases.

Acknowledgement

We would like to thank Asso.Prof.Kawee Pupaibul, the former head of department of Microbiology, Faculty of Medicine, Chulalongkorn University, for his guidance and encouragement throughout the study.

Grateful acknowledgement is made to Miss. Sumanee Sirilertpanrana for her excellent technical assistance. Many thanks to the Streptococcal Department, Statens Serum Institut, Denmark for providing us the reference strains of *Streptococcus pneumoniae*.

References

1. Christine MW, Khan FA, Khan A. Clinical in Chest Medicine. Philidelphia: WS Saunders, 1987. 393-4
2. Burr ML, Grays SJ, Howells CH. A survey of nasal *Streptococcus Pneumoniae* in children. J Hyg (Lond) 1982 Jun; 88(3) : 425-31
3. Holliday MG. Pneumococcal typing by polyvalent counter-immunoelectrophoresis. J Immunol Methods 1981; 46(2) : 243-9
4. Nanthapisud P, Ananchanachai V, Sukonthaman A. The serological types of *streptococcus pneumoniae* isolated from patients with systemic

- diseases. *Chula Med J* 1989 May; 33(5) : 375-80
5. Codoz M, Armand J, Arminjon, Michel JP, Michel M, Denis F, and Schiffman G. A new 23 valent pneumococcal vaccine : immunogenicity and pathogenicity in adults. *J Biol Scand* 1985; 13:261-5
 6. Coonrod JD, Rytel MW. Detection of type-specific pneumococcal antigens by counterimmunoelectrophoresis: methodology and immunologic properties of pneumococcal antigens. *J Lab Clin Med* 1973 May; 81(5) : 770-7
 7. Ratanavararak M, Vejijiva S. Serological Procedure and Interpretation. Bangkok: Chulalongkorn Press, 1984.28-32
 8. Severin WP. Latex agglutination in the diagnosis of Meningococcal Meningitis. *J Clin Pathol* 1972 Dec; 25(12) : 1079-82
 9. Newman RB, Stevens RW, Hagaufar A. Latex agglutination test for the diagnosis of Haemophilus Infuenzae meningitis. *J Lab Clin Med* 1970 Jan; 76(1) : 107-13
 10. Kaldor J, Asznovicz R, Buist GP. Latex agglutination in diagnosis of bacterial infections with special reference to patients with meningitis and septicemia. *Am J Clin Pathol* 1977 Aug; 68 (2) : 284-9
 11. Ajello GW, Bolan GA, Hayes PS, Lehmann D, Montgomery J, Feeley JC, Peilino CA, Broome CV. Commercial latex agglutination tests for detection of Haemophilus Infuenzae type B and Streptococcus Pneumoniae antigens in patients with bacteremic pneumonia. *J Clin Micro* 1987 Aug; 25(8) : 1388-9
 12. David WS, Ward JJ, Slber GR. Advantage of latex agglutination over countercurrent immunoelectrophoresis in the detection of Haemophilus Infuenzae type B antigen in serum. *Pediatrics* 1981 Dec; 68 (6) : 888-90
 13. Lund E. Laboratory diagnosis of pneumococcal infection. *Bull Org Mond Sante* 1960; 23(1) : 5-13
 14. Finegold SM. Bailey & Scott's Diagnostic Microbiology. 7th ed. St.Louis: CV. Mosby, 1986.
 15. Brown SL, Klein GC, McKinney FT, Jones WL. Safranin-O strained antigen microagglutination test for detection of brucella antibodies. *J Clin Micro* 1981 Feb; 13(2) : 398-400
 16. Goding JW. Conjugation of antibodies with fluorochromes modification to the standard methods. *J Immunol Methods* 1976; 13(3-4) : 215-26
 17. Lowry OH, Rosenbrough NJ, Farr AL, Randall RJ. Protein measurement with the folin phenol reagent. *J Biol Chem* 1951; 193 : 265-75
 18. Teresa L, Branefors P. Nature of cross-reactivity between Haemophilus infuenzae type A and B and Streptococcus pneumoniae type 6A and 6B. *Acta Pathol Microbiol Immunol Scand. Sect C* 1983; 91 : 371-6
 19. Richard AI, Anderson PW. Cross - reactivity with Escherichia coli K 100 in the Human Serum Anticapsular Antibody Response to Haemophilus infuenzae type B. *J Immunol* 1982 Mar; 128(3) : 1267-70